



World Class Accreditation

The American Association for Laboratory Accreditation

Accredited Laboratory

A2LA has accredited

CMM TECHNOLOGY, INC.

San Clemente, CA

for technical competence in the field of

Calibration

This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2005 *General Requirements for the Competence of Testing and Calibration Laboratories*. This laboratory also meets any additional program requirements in the field of calibration. This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (*refer to joint ISO-ILAC-IAF Communiqué dated 8 January 2009*).

Presented this 1st day of December 2009.



A handwritten signature in black ink, appearing to read "Peter Meyer", written over a horizontal line.

President & CEO
For the Accreditation Council
Certificate Number 2061.01
Valid to May 31, 2011

For the calibrations to which this accreditation applies, please refer to the laboratory's Calibration Scope of Accreditation.



SCOPE OF ACCREDITATION TO ISO/IEC 17025:2005

CMM TECHNOLOGY, INC.
 1230 Puerta del Sol
 San Clemente, CA 92673
 Lynne Helquist Phone: 800-648-6944

CALIBRATION

Valid To: May 31, 2011

Certificate Number: 2061.01

In recognition of the successful completion of the A2LA evaluation process, accreditation is granted to this laboratory to perform the following calibrations¹:

I. Dimensional

Parameter/Equipment	Range	CMC ^{2,4} (±)	Comments
Coordinate Measuring Machines (CMM) ³ –			Per ANSI B89.4.1-1997 with:
Repeatability	Sphere Diameter	12 µin	Sphere
Linear Accuracy	Up to 24 in (24 to 48) in (48 to 72) in (72 to 96) in (96 to 120) in	(14 + 13L) µin (330 + 4.8L) µin (470 + 11L) µin (570 + 14L) µin (660 + 17L) µin	Step gage
Volumetric Performance	Up to 240 in Ball Bar Length: (0 to 32) in	(11 + 2.1L) µin (26 + 5.5L) µin	Laser system Ball bar
Squareness	Ball Bar Length: (0 to 32) in	(26 + 5.5L) µin	Ball bar

¹ This laboratory offers commercial calibration service and field calibration service.

² Calibration and Measurement Capability (CMC) is the smallest uncertainty of measurement that a laboratory can achieve within its scope of accreditation when performing more or less routine calibrations of nearly ideal measurement standards or nearly ideal measuring equipment. Calibration and Measurement Capabilities represent expanded uncertainties expressed at approximately the 95 % level of confidence, usually using a coverage factor of $k = 2$. The actual measurement uncertainty of a specific calibration performed by the laboratory may be greater than the CMC due to the behavior of the customer's device and to influences from the circumstances of the specific calibration.

³ Field calibration service is available for this calibration and this laboratory meets A2LA R104 – *General Requirements: Accreditation of Field Testing and Field Calibration Laboratories* for these calibrations. Please note the actual measurement uncertainties achievable on a customer's site can normally be expected to be larger than the CMC found on the A2LA Scope. Allowance must be made for aspects such as the environment at the place of calibration and for other possible adverse effects such as those caused by transportation of the calibration equipment. The usual allowance for the actual uncertainty introduced by the item being calibrated, (e.g. resolution) must also be considered and this, on its own, could result in the actual measurement uncertainty achievable on a customer's site being larger than the CMC.

⁴ In the statement of CMC, L is the numerical value of the nominal length of the device measured in inches.

